



Customer Notification

DMT1/PM205 Pressure Calibrator



Announcement date:	16-September-2013
Effective date:	16-September-2013
Document number:	PS00138-IKM
User manual:	Rev. 01.s009, page 1 and 8
Notification category:	General notice
Subject:	Expected pressure inaccuracy after exceeding recommended calibration interval.

The recommended calibration/service interval for the DMT-1/PM205 Pressure calibrator is 12 months. By exceeding this interval it should be notified that the typical inaccuracy/drift of the DMT-1/PM205 will be $\pm 0,5\%$ FS per year. This estimate is limited for a periode of 3 years

Paul Egil Borvik Larsen
HSEQ Manager

Bjørn Sverre Mikkelsen
Manager Export Sales and
Calibration Lab.